



**Statement of Accuracy**  
**Cyanide Vacu-vials® Kit, Cat. No. K-3803**

This statement of accuracy is based on laboratory tests performed under ideal testing conditions using standards of known concentration prepared in deionized water.

Instrument Platform	Test Range, ppm	Accuracy (expressed as ppm at concentrations across the test range)			
		0 ppm	0.040 ppm	0.100 ppm	0.300 ppm
Spectrophotometer	0 - 0.400	≤ 0.010	+/- 0.016	+/- 0.030	+/- 0.060
V-2000	0 - 0.400	≤ 0.020	+/- 0.016	+/- 0.040	+/- 0.090
V-3000 Series	0 - 0.400	≤ 0.020	+/- 0.016	+/- 0.040	+/- 0.090

Updated 12/01/2020