

Nitrite CHEMets® Kit

K-7006/R-7006: 0 - 0.1 & 0 - 1 ppm N

Safety Information

Read SDS (available at www.chemetrics.com) before performing this test procedure. Wear safety glasses and protective gloves.

Test Procedure

1. Add 5 drops of A-7004 Acidifier Solution to the empty sample cup (fig. 1).
2. Fill the sample cup to the 25 mL mark with the sample to be tested (fig. 2). Stir to mix the contents of the cup.
3. Place the CHEMet ampoule, tip first, into the sample cup. Snap the tip. The ampoule will fill leaving a bubble for mixing (fig. 3).
4. To mix the ampoule, invert it several times, allowing the bubble to travel from end to end.
5. Dry the ampoule. Obtain a test result **8 minutes** after snapping the tip.
6. Obtain a test result using the appropriate comparator.

a. Low Range Comparator (fig. 4):

Place the ampoule, flat end first, into the comparator. Hold the comparator up toward a source of light and view from the bottom. Rotate the comparator until the best color match is found.




Figure 1




Figure 2




Figure 3



Figure 4

- b. **High Range Comparator (fig. 5):** Place the ampoule between the color standards until the best color match is found.




Figure 5

Test Method

The Nitrite CHEMets®¹ test kit employs the azo dye method.^{2,3} In an acidic solution, nitrite diazotizes with the primary aromatic amine N-(1-naphthyl)ethylenediamine dihydrochloride (NED) and then couples with sulfanilic acid to produce a highly colored azo dye. The resulting pink color is proportional to the nitrite concentration in the sample.

1. CHEMets is a registered trademark of CHEMetrics, Inc. U.S. Patent No. 3,634,038
2. APHA Standard Methods, 23rd ed., Method 4500-NO₂⁻ B - 2000.
3. EPA Methods for Chemical Analysis of Water and Wastes, Method 354.1 (1983).

Visit www.chemetrics.com to view product demonstration videos.
Always follow the test procedure above to perform a test.



www.chemetrics.com
4295 Catlett Road, Midland, VA 22728 U.S.A.
Phone: (800) 356-3072; Fax: (540) 788-4856
E-Mail: orders@chemetrics.com

July 22, Rev. 1